

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL012622\
 Data File : VL038292.D
 Acq On : 26 Jan 2022 17:25
 Operator : SY/AP
 Sample : N1268-02DL 10X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 OA-1DL

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 01/27/2022
 Supervised By : Mahesh Dadoda 01/28/2022

Quant Time: Jan 27 04:06:15 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL012422A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Mon Jan 24 14:12:48 2022
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.64	49	797291	10.00	ppbv	-0.04
33) 1,4-Difluorobenzene	7.14	114	2115742	10.00	ppbv	-0.04
55) Chlorobenzene-d5	12.05	117	1705119	10.00	ppbv	-0.04

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.38	95	1280777	9.90	ppbv	-0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.60	45	17153m	3.092	ppbv	
15) Acetone	3.86	43	21483m	0.306	ppbv	
19) Isopropyl Alcohol	3.98	45	6618m	0.151	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL012622\
Data File : VL038292.D
Acq On : 26 Jan 2022 17:25
Operator : SY/AP
Sample : N1268-02DL 10X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
OA-1DL

Quant Time: Jan 27 04:06:15 2022
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL012422
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Mon Jan 24 14:12:48 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 01/27/2022
Supervised By :Mahesh Dadoda 01/28/2022

