

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL012720\
 Data File : VL034558.D
 Acq On : 27 Jan 2020 12:20
 Operator : SY/AP
 Sample : VL0127ABL01
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0127ABL01

Manual Integrations
 APPROVED

MMDadoda
 1/28/2020 2:02:46 PM

Quant Time: Jan 28 00:43:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL123019AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jan 0
 QLast Update : Thu Jan 09 03:01:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	945449	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.19	114	1978253	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.13	117	1955588	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.46	95	1634221	10.01	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.85	43	43520	0.315	ppbv	# 72
20) Methylene Chloride	4.33	84	12472m	0.230	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL012720\
Data File : VL034558.D
Acq On : 27 Jan 2020 12:20
Operator : SY/AP
Sample : VL0127ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0127ABL01

Manual Integrations
APPROVED

MMDadoda
1/28/2020 2:02:46 PM

Quant Time: Jan 28 00:43:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL123019AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jan
QLast Update : Thu Jan 09 03:01:19 2020
Response via : Initial Calibration

