

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL012720\
 Data File : VL034565.D
 Acq On : 27 Jan 2020 17:41
 Operator : SY/AP
 Sample : OC012020 CAN 10300
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC012020 CAN 10300

Quant Time: Jan 28 12:08:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL123019AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jan 09 03:01:19 2020
 QLast Update : Thu Jan 09 03:01:19 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	966117	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.20	114	1952070	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.13	117	1908484	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.47	95	1601539	10.06	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.60%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL012720\
Data File : VL034565.D
Acq On : 27 Jan 2020 17:41
Operator : SY/AP
Sample : OC012020 CAN 10300
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC012020 CAN 10300

Quant Time: Jan 28 12:08:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL123019AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jan 09 03:01:19 2020
QLast Update : Thu Jan 09 03:01:19 2020
Response via : Initial Calibration

