

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL012722\
 Data File : VL038309.D
 Acq On : 27 Jan 2022 23:11
 Operator : SY/AP
 Sample : N1300-02DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA-20DL

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 01/28/2022
 Supervised By : Mahesh Dadoda 01/28/2022

Quant Time: Jan 28 06:58:01 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL012422A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Mon Jan 24 14:12:48 2022
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	776892	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.16	114	1984865	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.06	117	1668443	10.00	ppbv	-0.03

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.39	95	1242992	9.82	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.20%

Target Compounds

	R.T.	QIon	Response	Conc	Units	Qvalue
9) Propene	3.02	41	9216	0.183	ppbv	85
13) Ethanol	3.59	45	630233	116.606	ppbv	99
15) Acetone	3.85	43	52629m	0.768	ppbv	
19) Isopropyl Alcohol	3.97	45	29374m	0.686	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL012722\
Data File : VL038309.D
Acq On : 27 Jan 2022 23:11
Operator : SY/AP
Sample : N1300-02DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA-20DL

Quant Time: Jan 28 06:58:01 2022
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL012422
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Mon Jan 24 14:12:48 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 01/28/2022
Supervised By :Mahesh Dadoda 01/28/2022

