

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL012722\
 Data File : VL038310.D
 Acq On : 27 Jan 2022 23:48
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 01/28/2022
 Supervised By : Mahesh Dadoda 01/28/2022

Quant Time: Jan 28 07:00:26 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL012422A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Mon Jan 24 14:12:48 2022
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	703878	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.16	114	1947925	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.06	117	1603948	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.39	95	1203580	9.89	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.90%

Target Compounds					Qvalue
13) Ethanol	3.62	45	4590m	0.937	ppbv

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL012722\
Data File : VL038310.D
Acq On : 27 Jan 2022 23:48
Operator : SY/AP
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

Quant Time: Jan 28 07:00:26 2022

Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL012422

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:

QLast Update : Mon Jan 24 14:12:48 2022

Response via : Initial Calibration

Reviewed By :Semsettin Yesilyurt 01/28/2022
Supervised By :Mahesh Dadoda 01/28/2022

