

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL012722\
 Data File : VL038314.D
 Acq On : 28 Jan 2022 9:10
 Operator : SY/AP
 Sample : QC012722 CAN 10667
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC012722 CAN 10667

Quant Time: Jan 29 06:05:10 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL012422AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Jan 24 14:12:48 2022
 QLast Update : Mon Jan 24 14:12:48 2022
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	713526	10.00	ppbv	-0.02
33) 1,4-Difluorobenzene	7.16	114	1885311	10.00	ppbv	-0.02
55) Chlorobenzene-d5	12.07	117	1576696	10.00	ppbv	-0.02

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.40	95	1198265	10.01	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.10%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL012722\
Data File : VL038314.D
Acq On : 28 Jan 2022 9:10
Operator : SY/AP
Sample : QC012722 CAN 10667
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC012722 CAN 10667

Quant Time: Jan 29 06:05:10 2022
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL012422AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Jan 24 14:12:48 2022
QLast Update : Mon Jan 24 14:12:48 2022
Response via : Initial Calibration

