

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL012819\
 Data File : VL033118.D
 Acq On : 28 Jan 2019 18:13
 Operator : SY/AP
 Sample : VSTDIC0.1
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.1

Manual Integrations
 APPROVED

apatel
 1/29/2019 2:36:17 PM

Quant Time: Jan 28 19:54:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL012819AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Jan 28 2019
 QLast Update : Mon Jan 28 16:32:44 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	1137162	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.28	114	3775443	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	3670850	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	2802664	9.96	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.29	62	9121	0.122	ppbv	99
32) 1,1,1-Trichloroethane	6.62	97	34176	0.158	ppbv	94
35) Carbon Tetrachloride	7.13	117	34487m	0.106	ppbv	
38) Trichloroethene	7.96	130	12809m	0.101	ppbv	
52) Tetrachloroethene	11.37	164	8349m	0.076	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.83	83	33728m	0.108	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL012819\
Data File : VL033118.D
Acq On : 28 Jan 2019 18:13
Operator : SY/AP
Sample : VSTDIC0.1
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.1

Manual Integrations
APPROVED

apatel
1/29/2019 2:36:17 PM

Quant Time: Jan 28 19:54:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL012819AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Jan
QLast Update : Mon Jan 28 16:32:44 2019
Response via : Initial Calibration

