

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL012819\
 Data File : VL033127.D
 Acq On : 29 Jan 2019 4:25
 Operator : SY/AP
 Sample : K1245-01DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 S-01DL

Manual Integrations
 APPROVED

apatel
 1/29/2019 2:38:37 PM

Quant Time: Jan 29 07:02:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL012819AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Jan 29 2019
 QLast Update : Tue Jan 29 06:21:00 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	1181249	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	2974288	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	2558475	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	1886365	9.75	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	3.05	41	14683	0.241	ppbv #	79
13) Ethanol	3.65	45	295872	13.510	ppbv	99
15) Acetone	3.92	43	85752	0.538	ppbv	100
20) Methylene Chloride	4.41	84	25337	0.350	ppbv	94
26) Hexane	5.78	57	19050	0.140	ppbv #	61
29) Chloroform	5.84	83	43247m	0.193	ppbv	
52) Tetrachloroethene	11.37	164	17498	0.200	ppbv	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL012819\
Data File : VL033127.D
Acq On : 29 Jan 2019 4:25
Operator : SY/AP
Sample : K1245-01DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
S-01DL

Manual Integrations
APPROVED

apatel
1/29/2019 2:38:37 PM

Quant Time: Jan 29 07:02:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL012819AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Jan
QLast Update : Tue Jan 29 06:21:00 2019
Response via : Initial Calibration

