

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL012821\
 Data File : VL036273.D
 Acq On : 28 Jan 2021 15:27
 Operator : SY/AP
 Sample : VL0128ABL01
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0128ABL01

Manual Integrations
 APPROVED

MMDadoda
 2/1/2021 1:52:01 PM

Quant Time: Jan 29 02:52:11 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL011321AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jan 14 16:13:38 2021
 QLast Update : Thu Jan 14 16:13:38 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1648239	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	3486717	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.08	117	3210416	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.41	95	2653659	9.99	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.90%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.87	43	44629m	0.221	ppbv	
20) Methylene Chloride	4.34	84	8556m	0.133	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL012821\
Data File : VL036273.D
Acq On : 28 Jan 2021 15:27
Operator : SY/AP
Sample : VL0128ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0128ABL01

Manual Integrations
APPROVED

MMDadoda
2/1/2021 1:52:01 PM

Quant Time: Jan 29 02:52:11 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL011321AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jan
QLast Update : Thu Jan 14 16:13:38 2021
Response via : Initial Calibration

