

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL012821\
 Data File : VL036288.D
 Acq On : 29 Jan 2021 8:49
 Operator : SY/AP
 Sample : M1245-02DL 80X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SS-2DL

Manual Integrations
 APPROVED

MMDadoda
 2/1/2021 1:53:59 PM

Quant Time: Jan 31 17:20:54 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL011321AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jan 14 16:13:38 2021
 QLast Update : Thu Jan 14 16:13:38 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1523958	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	3403977	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.08	117	3100815	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.41	95	2607669	10.16	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.60	45	1058559m	75.638	ppbv	
15) Acetone	3.84	43	1681317	9.008	ppbv	99
17) tert-Butyl alcohol	4.29	59	1173444m	7.619	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL012821\
Data File : VL036288.D
Acq On : 29 Jan 2021 8:49
Operator : SY/AP
Sample : M1245-02DL 80X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SS-2DL

Manual Integrations
APPROVED

MMDadoda
2/1/2021 1:53:59 PM

Quant Time: Jan 31 17:20:54 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL011321AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jan
QLast Update : Thu Jan 14 16:13:38 2021
Response via : Initial Calibration

