

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL013023\
 Data File : VL040131.D
 Acq On : 30 Jan 2023 14:01
 Operator : SY/MD
 Sample : QC012423 CAN 10652
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC012423 CAN 10652

Quant Time: Jan 31 05:12:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL011623AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Mon Jan 16 14:40:47 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.182	49	912601	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.642	114	2449056	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.461	117	2149585	10.000	ppbv	-0.01
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.773	95	1707648	9.558	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	95.600%
Target Compounds						Qvalue
13) Ethanol	3.217	45	105	0.025	ppbv	# 38

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL013023\
Data File : VL040131.D
Acq On : 30 Jan 2023 14:01
Operator : SY/MD
Sample : QC012423 CAN 10652
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC012423 CAN 10652

Quant Time: Jan 31 05:12:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL011623AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Mon Jan 16 14:40:47 2023
Response via : Initial Calibration

