


Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL020322\
 Data File : VL038326.D
 Acq On : 3 Feb 2022 8:42
 Operator : SY/AP
 Sample : BFB
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL012422AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Mon Jan 24 14:12:48 2022
 Last Update : Mon Jan 24 14:12:48 2022



AutoFind: Scans 3607, 3608, 3609; Background Corrected with Scan 3573

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	22.1	148032	PASS
75	95	30	66	53.9	360277	PASS
95	95	100	100	100.0	668757	PASS
96	95	5	9	6.8	45208	PASS
173	174	0.00	2	1.2	5052	PASS
174	95	50	120	62.4	417493	PASS
175	174	4	9	7.4	30688	PASS
176	174	93	101	96.4	402517	PASS
177	176	5	9	6.7	26768	PASS