

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL020420\
 Data File : VL034590.D
 Acq On : 4 Feb 2020 21:51
 Operator : SY/AP
 Sample : L1357-03DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA-1DL

Manual Integrations
 APPROVED

MMDadoda
 2/5/2020 2:50:50 PM

Quant Time: Feb 05 06:59:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL013020AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jan 3
 QLast Update : Thu Jan 30 17:16:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	932339	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.19	114	1934123	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.12	117	1980203	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.46	95	1689087	10.09	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	100.90%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	2.99	41	38132m	0.869	ppbv	
13) Ethanol	3.60	45	6609m	0.563	ppbv	
15) Acetone	3.84	43	55790	0.407	ppbv #	89
20) Methylene Chloride	4.34	84	19134m	0.367	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL020420\
Data File : VL034590.D
Acq On : 4 Feb 2020 21:51
Operator : SY/AP
Sample : L1357-03DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA-1DL

Manual Integrations
APPROVED

MMDadoda
2/5/2020 2:50:50 PM

Quant Time: Feb 05 06:59:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL013020AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jan
QLast Update : Thu Jan 30 17:16:44 2020
Response via : Initial Calibration

