

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL020420\
 Data File : VL034597.D
 Acq On : 5 Feb 2020 2:27
 Operator : SY/AP
 Sample : L1391-01DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV-1DL

Manual Integrations
 APPROVED

MMDadoda
 2/5/2020 2:51:24 PM

Quant Time: Feb 05 08:03:03 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL013020AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jan 30 2020
 QLast Update : Thu Jan 30 17:16:44 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	980656	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.19	114	1993923	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.12	117	2027036	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.45	95	1732347	10.11	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	2.98	41	32643m	0.708	ppbv	
13) Ethanol	3.59	45	16694m	1.351	ppbv	
15) Acetone	3.84	43	81924	0.568	ppbv	92
20) Methylene Chloride	4.33	84	22302	0.406	ppbv	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL020420\
Data File : VL034597.D
Acq On : 5 Feb 2020 2:27
Operator : SY/AP
Sample : L1391-01DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV-1DL

Manual Integrations
APPROVED

MMDadoda
2/5/2020 2:51:24 PM

Quant Time: Feb 05 08:03:03 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL013020AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Thu Jan
QLast Update : Thu Jan 30 17:16:44 2020
Response via : Initial Calibration

