

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL020519\
 Data File : VL033173.D
 Acq On : 6 Feb 2019 4:24
 Operator : SY/AP
 Sample : VSTDIC0.1
 Misc : 400ml/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.1

Manual Integrations
 APPROVED

MMDadoda
 2/20/2019 12:21:58 PM

Quant Time: Feb 20 05:56:59 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL020519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Feb 20 05:56:59 2019
 QLast Update : Wed Feb 20 05:04:48 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	968950	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	2608690	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	2587316	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	2065194	10.24	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
5) Vinyl Chloride	3.30	62	7078m	0.110	ppbv	
32) 1,1,1-Trichloroethane	6.62	97	19372m	0.100	ppbv	
35) Carbon Tetrachloride	7.13	117	19123	0.096	ppbv	# 78
38) Trichloroethene	7.96	130	12890m	0.114	ppbv	
52) Tetrachloroethene	11.38	164	11024m	0.109	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.84	83	27461m	0.102	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL020519\
Data File : VL033173.D
Acq On : 6 Feb 2019 4:24
Operator : SY/AP
Sample : VSTDIC0.1
Misc : 400ml/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.1

Manual Integrations
APPROVED

MMDadoda
2/20/2019 12:21:58 PM

Quant Time: Feb 20 05:56:59 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL020519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Feb
QLast Update : Wed Feb 20 05:04:48 2019
Response via : Initial Calibration

