

Data Path : W:\HPCHEM1\MSVOA_L\DATA\VL020916\
 Data File : VL026824.D
 Acq On : 9 Feb 2016 17:35
 Operator : SY/AP
 Sample : VSTDIC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC002

Quant Time: Feb 10 02:08:38 2016
 Quant Method : W:\HPCHEM1\MSVOA_L\METHODS\VL020916AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Feb 10 02:03:34 2016
 QLast Update : Wed Feb 10 02:03:34 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	6.69	49	1694782	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	8.37	114	4564495	10.00	ppbv	-0.01
55) Chlorobenzene-d5	13.79	117	4469471	10.00	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	16.33	95	3177756	8.38	ppbv	-0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	83.80%

Target Compounds						Qvalue
2) Dichlorodifluoromethane	3.67	85	771912	1.43	ppbv	99
3) Chlorodifluoromethane	3.60	51	598108	1.84	ppbv	99
4) Chloromethane	3.78	50	226543	2.25	ppbv	99
5) Vinyl Chloride	3.92	62	242277	2.03	ppbv	97
6) Bromomethane	4.17	94	205978	2.00	ppbv	99
7) Chloroethane	4.28	64	110115	2.01	ppbv	95
8) Dichlorotetrafluoroethane	3.84	85	662492	1.68	ppbv	99
9) Propene	3.63	41	156191	2.34	ppbv	100
10) Heptane	9.40	43	501025	1.96	ppbv	98
11) Trichlorofluoromethane	4.73	101	890271	1.44	ppbv	98
12) 1,1,2-Trichlorotrifluoroet	5.34	101	618183	1.70	ppbv	99
13) Ethanol	4.38	45	28838	1.88	ppbv	94
14) Bromoethene	4.49	108	228962	1.95	ppbv	100
15) Acetone	4.64	43	527976	1.32	ppbv #	76
16) 1,3-Butadiene	4.00	39	188317	1.73	ppbv	98
17) tert-Butyl alcohol	5.18	59	260637	1.22	ppbv	95
18) 1,1-Dichloroethene	5.12	96	263201	1.88	ppbv	96
19) Isopropyl Alcohol	4.81	45	163982	1.56	ppbv #	95
20) Methylene Chloride	5.18	84	239224	1.85	ppbv	98
21) Allyl Chloride	5.26	41	285019	1.69	ppbv	99
22) trans-1,2-Dichloroethene	5.81	96	234697	1.76	ppbv	97
23) Vinyl Acetate	6.02	43	670643	1.58	ppbv	98
24) 1,1-Dichloroethane	5.95	63	504098	1.69	ppbv	99
25) Ethyl Acetate	6.71	43	764019	1.95	ppbv	99
26) Hexane	6.70	57	404156	2.07	ppbv	98
27) Carbon Disulfide	5.43	76	691528	1.78	ppbv	99
28) Methyl tert-Butyl Ether	6.00	73	645973	1.48	ppbv	100
29) Chloroform	6.79	83	608513	1.39	ppbv	99
30) Cyclohexane	8.32	84	328928	1.63	ppbv	87
31) cis-1,2-Dichloroethene	6.56	61	332078	1.56	ppbv	99
32) 1,1,1-Trichloroethane	7.63	97	603598	1.21	ppbv	100
34) 2-Butanone	6.24	43	466773	2.20	ppbv	98
35) Carbon Tetrachloride	8.20	117	646799	1.43	ppbv	98
36) Benzene	8.06	78	737666	1.93	ppbv	99
37) 1,2-Dichloroethane	7.41	62	453037	1.38	ppbv	100
38) Trichloroethene	9.11	130	329237	1.82	ppbv	98
39) 1,2-Dichloropropane	8.87	63	273942	2.13	ppbv	99
40) 1,4-Dioxane	9.20	88	53053	1.69	ppbv #	1
41) Tetrahydrofuran	7.18	42	228677	2.38	ppbv	99
42) Bromodichloromethane	9.07	83	609722	1.43	ppbv	97
43) Methyl Methacrylate	9.31	69	256454	1.74	ppbv	99

Data Path : W:\HPCHEM1\MSVOA_L\DATA\VL020916\
 Data File : VL026824.D
 Acq On : 9 Feb 2016 17:35
 Operator : SY/AP
 Sample : VSTDIC002
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC002

Quant Time: Feb 10 02:08:38 2016
 Quant Method : W:\HPCHEM1\MSVOA_L\METHODS\VL020916AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Feb 10 02:03:34 2016
 QLast Update : Wed Feb 10 02:03:34 2016
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,2,4-Trimethylpentane	9.13	57	1306104	2.15	ppbv	99
45) t-1,3-Dichloropropene	10.69	75	335677	1.32	ppbv	99
46) cis-1,3-Dichloropropene	10.06	75	417660	1.53	ppbv	99
47) 1,1,2-Trichloroethane	10.92	97	295716	1.45	ppbv	99
48) Dibromochloromethane	11.85	129	478982	1.25	ppbv	100
49) Bromoform	14.86	173	349775	1.06	ppbv	99
50) 4-Methyl-2-Pentanone	10.14	43	552136	1.87	ppbv	99
51) 2-Hexanone	11.68	43	458462	1.61	ppbv #	99
52) Tetrachloroethene	12.84	164	279026	1.30	ppbv	99
53) Toluene	11.28	91	845070	1.50	ppbv	99
54) 1,2-Dibromoethane	12.19	107	444235	1.29	ppbv	99
56) 1,1,1,2-Tetrachloroethane	13.83	131	370079	1.57	ppbv #	1
57) Chlorobenzene	13.86	112	702313	1.55	ppbv	98
58) Ethyl Benzene	14.46	91	1120180	1.53	ppbv	96
59) m/p-Xylene	14.76	91	985053	1.54	ppbv #	41
60) o-Xylene	15.54	91	994282	1.40	ppbv	98
61) Styrene	15.36	104	342450	1.34	ppbv	99
62) Isopropylbenzene	16.59	105	1210981	1.31	ppbv	98
63) 1,1,2,2-Tetrachloroethane	15.53	83	647479	1.42	ppbv	100
64) n-propylbenzene	17.57	120	302101	1.17	ppbv	82
65) tert-Butylbenzene	18.86	119	1088256	1.16	ppbv	99
66) Benzyl Chloride	19.16	91	320386	0.91	ppbv	98
67) sec-Butylbenzene	19.46	105	1467983	1.07	ppbv	98
69) p-Isopropyltoluene	19.84	119	1178447	1.07	ppbv	99
70) n-Butylbenzene	20.80	91	1149510	1.00	ppbv	97
71) 2-Chlorotoluene	17.47	91	844009	1.17	ppbv	100
72) 4-Ethyltoluene	17.87	105	943701	1.17	ppbv	99
73) 1,3,5-Trimethylbenzene	18.03	105	894797	1.13	ppbv	94
74) 1,2,4-Trimethylbenzene	18.88	105	1002549	1.14	ppbv	97
75) 1,3-Dichlorobenzene	19.17	146	597522	1.04	ppbv	99
76) 1,4-Dichlorobenzene	19.32	146	580887	1.01	ppbv	99
77) 1,2-Dichlorobenzene	20.08	146	568237	0.95	ppbv	100
78) Hexachloro-1,3-Butadiene	24.72	225	277395	0.95	ppbv	99
79) Naphthalene	23.98	128	530930	0.70	ppbv	99
80) Naphthalene,2-methyl-	26.09	142	210417	1.60	ppbv	97
81) 1,2,4-Trichlorobenzene	23.78	180	271057	0.81	ppbv	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA_L\DATA\VL020916\
Data File : VL026824.D
Acq On : 9 Feb 2016 17:35
Operator : SY/AP
Sample : VSTDIC002
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC002

Quant Time: Feb 10 02:08:38 2016
Quant Method : W:\HPCHEM1\MSVOA_L\METHODS\VL020916AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Feb 10 02:03:34 2016
QLast Update : Wed Feb 10 02:03:34 2016
Response via : Initial Calibration





#1

Bromochloromethane

Concen: 10.00 ppbv

RT: 6.69 min Scan# 591

Delta R.T. -0.01 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC002


Tgt Ion: 49 Resp: 1694782

Ion Ratio Lower Upper

49 100

128 58.4 29.7 89.1




130 75.3 38.3 114.8



Abundance Ion 49.10 (48.80 to 49.80): VL0

Ion 128.00 (127.70 to 128.70): VL0

Ion 130.00 (129.70 to 130.70): VL0



#2

Dichlorodifluoromethane

Concen: 1.43 ppbv

RT: 3.67 min Scan# 96

Delta R.T. -0.00 min

Lab File: VL026824.D


Acq: 9 Feb 2016 17:35

Tgt Ion: 85 Resp: 771912

Ion Ratio Lower Upper



85 100


87 33.3 27.3 40.9



Abundance Ion 85.00 (84.70 to 85.70): VL0

Ion 87.00 (86.70 to 87.70): VL0





#3

Chlorodifluoromethane

Concen: 1.84 ppbv

RT: 3.60 min Scan# 84

Delta R.T. -0.00 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35

Instrument :

MSVOA_L

ClientSampleId :


VSTDIC002

Tgt Ion: 51 Resp: 598108

Ion Ratio Lower Upper




51 100

67 20.0 16.4 24.6



Abundance Ion 51.00 (50.70 to 51.70): VL0

Ion 67.00 (66.70 to 67.70): VL0



#4

Chloromethane

Concen: 2.25 ppbv

RT: 3.78 min Scan# 114

Delta R.T. -0.00 min

Lab File: VL026824.D


Acq: 9 Feb 2016 17:35

Tgt Ion: 50 Resp: 226543

Ion Ratio Lower Upper



50 100


52 32.4 26.3 39.5



Abundance Ion 50.10 (49.80 to 50.80): VL0

Ion 52.10 (51.80 to 52.80): VL0





#5

Vinyl Chloride

Concen: 2.03 ppbv

RT: 3.92 min Scan# 137

Delta R.T. 0.01 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35

Instrument :

MSVOA_L

ClientSampleId :





VSTDIC002

Tgt Ion: 62 Resp: 242277

Ion Ratio Lower Upper

62 100

64 30.5 25.8 38.6



#6

Bromomethane

Concen: 2.00 ppbv

RT: 4.17 min Scan# 178

Delta R.T. -0.00 min

Lab File: VL026824.D




Acq: 9 Feb 2016 17:35


Tgt Ion: 94 Resp: 205978

Ion Ratio Lower Upper

94 100

96 92.6 73.2 109.8





#7

Chloroethane

Concen: 2.01 ppbv

RT: 4.28 min Scan# 195

Delta R.T. 0.01 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35

Instrument :

MSVOA_L

ClientSampleId :


VSTDIC002

Tgt Ion: 64 Resp: 110115

Ion Ratio Lower Upper

64 100

66 34.4 25.4 38.0



Abundance Ion 64.10 (63.80 to 64.80): VL0

Ion 66.10 (65.80 to 66.80): VL0

4.28

60000

50000

40000



30000

20000

10000

0

Time-->



#8

Dichlorotetrafluoroethane

Concen: 1.68 ppbv

RT: 3.84 min Scan# 123

Delta R.T. -0.00 min

Lab File: VL026824.D


Acq: 9 Feb 2016 17:35

Tgt Ion: 85 Resp: 662492

Ion Ratio Lower Upper

85 100

135 78.2 63.2 94.8



Abundance Ion 85.00 (84.70 to 85.70): VL0

Ion 135.00 (134.70 to 135.70): V

3.84

400000


300000


200000

100000

0

Time-->





#9

Propene

Concen: 2.34 ppbv

RT: 3.63 min Scan# 89

Delta R.T. 0.01 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35

Instrument :

MSVOA_L

ClientSampleId :


VSTDIC002

Tgt Ion: 41 Resp: 156191

Ion Ratio Lower Upper




41 100

42 66.1 52.8 79.2



Abundance Ion 41.10 (40.80 to 41.80): VL0

Ion 42.10 (41.80 to 42.80): VL0



#10

Heptane

Concen: 1.96 ppbv

RT: 9.40 min Scan# 1035

Delta R.T. -0.01 min

Lab File: VL026824.D


Acq: 9 Feb 2016 17:35

Tgt Ion: 43 Resp: 501025

Ion Ratio Lower Upper



43 100


57 56.3 44.0 66.0



Abundance Ion 43.10 (42.80 to 43.80): VL0

Ion 57.10 (56.80 to 57.80): VL0









#11
 Trichlorofluoromethane
 Concen: 1.44 ppbv
 RT: 4.73 min Scan# 269
 Delta R.T. -0.00 min
 Lab File: VL026824.D
 Acq: 9 Feb 2016 17:35




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC002


Tgt Ion:101 Resp: 890271
 Ion Ratio Lower Upper
 101 100
 103 64.6 52.7 79.1



#12
 1,1,2-Trichlorotrifluoroethane
 Concen: 1.70 ppbv
 RT: 5.34 min Scan# 370
 Delta R.T. -0.00 min
 Lab File: VL026824.D
 Acq: 9 Feb 2016 17:35

Tgt Ion:101 Resp: 618183
 Ion Ratio Lower Upper
 101 100
 151 85.7 69.4 104.0





#13

Ethanol

Concen: 1.88 ppbv

RT: 4.38 min Scan# 212

Delta R.T. 0.02 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35

Instrument :

MSVOA_L

ClientSampleId :





VSTDIC002

Tgt Ion: 45 Resp: 28838

Ion Ratio Lower Upper

45 100

46 43.5 15.8 63.4



#14

Bromoethene

Concen: 1.95 ppbv

RT: 4.49 min Scan# 230

Delta R.T. 0.01 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35




Tgt Ion: 108 Resp: 228962


Ion Ratio Lower Upper

108 100

106 106.0 84.6 126.8

79 16.6 13.4 20.0





#15

Acetone

Concen: 1.32 ppbv

RT: 4.64 min Scan# 255

Delta R.T. 0.01 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35

Instrument :

MSVOA_L

ClientSampleId :





VSTDIC002

Tgt Ion: 43 Resp: 527976

Ion Ratio Lower Upper

43 100

58 44.6 25.2 37.8#



#16

1,3-Butadiene

Concen: 1.73 ppbv

RT: 4.00 min Scan# 150

Delta R.T. -0.00 min

Lab File: VL026824.D




Acq: 9 Feb 2016 17:35


Tgt Ion: 39 Resp: 188317

Ion Ratio Lower Upper

39 100

54 91.8 74.7 112.1





#17

tert-Butyl alcohol

Concen: 1.22 ppbv

RT: 5.18 min Scan# 344

Delta R.T. 0.02 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35

Instrument :

MSVOA_L

ClientSampleId :


VSTDIC002

Tgt Ion: 59 Resp: 260637

Ion Ratio Lower Upper

59 100

57 12.6 8.5 12.7



Abundance Ion 59.10 (58.80 to 59.80): VL0

Ion 57.10 (56.80 to 57.80): VL0

5.18

100000

80000

60000



40000

20000

0

5.10 5.15 5.20 5.25 5.30

Time-->



#18

1,1-Dichloroethene

Concen: 1.88 ppbv

RT: 5.12 min Scan# 334

Delta R.T. -0.00 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35


Tgt Ion: 96 Resp: 263201

Ion Ratio Lower Upper

96 100

61 165.2 136.7 205.1

98 63.3 52.1 78.1



Abundance Ion 96.00 (95.70 to 96.70): VL0

Ion 61.10 (60.80 to 61.80): VL0

Ion 98.00 (97.70 to 98.70): VL0

300000

250000

200000

150000


100000


50000

0

5.05 5.10 5.15 5.20

Time-->





#19

Isopropyl Alcohol

Concen: 1.56 ppbv

RT: 4.81 min Scan# 282

Delta R.T. 0.02 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35

Instrument :

MSVOA_L

ClientSampleId :





VSTDIC002

Tgt Ion: 45 Resp: 163982

Ion Ratio Lower Upper

45 100

58 4.4 2.2 3.2#



#20

Methylene Chloride

Concen: 1.85 ppbv

RT: 5.18 min Scan# 344

Delta R.T. -0.00 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35

Tgt Ion: 84 Resp: 239224




Ion Ratio Lower Upper


84 100

49 120.5 94.9 142.3

51 38.8 29.5 44.3

86 64.2 51.9 77.9





#21

Allyl Chloride

Concen: 1.69 ppbv

RT: 5.26 min Scan# 357

Delta R.T. -0.00 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC002


Tgt Ion: 41 Resp: 285019

Ion Ratio Lower Upper

41 100

76 41.1 32.0 48.0




39 81.9 64.6 96.8



Abundance Ion 41.10 (40.80 to 41.80): VL0

Ion 76.10 (75.80 to 76.80): VL0

Ion 39.10 (38.80 to 39.80): VL0



#22

trans-1,2-Dichloroethene

Concen: 1.76 ppbv

RT: 5.81 min Scan# 446

Delta R.T. -0.00 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35


Tgt Ion: 96 Resp: 234697

Ion Ratio Lower Upper

96 100

61 148.7 121.6 182.4



98 61.0 50.6 76.0




Abundance Ion 96.00 (95.70 to 96.70): VL0

Ion 61.10 (60.80 to 61.80): VL0

Ion 98.00 (97.70 to 98.70): VL0





#23

Vinyl Acetate

Concen: 1.58 ppbv

RT: 6.02 min Scan# 481

Delta R.T. -0.01 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC002

Tgt Ion: 43 Resp: 670643


Ion Ratio Lower Upper

43 100

86 8.4 7.7 11.5

42 9.6 7.8 11.8

44 4.5 3.9 5.9






Abundance Ion 43.10 (42.80 to 43.80): VL0

Ion 86.10 (85.80 to 86.80): VL0

Ion 42.10 (41.80 to 42.80): VL0

Ion 44.10 (43.80 to 44.80): VL0



#24

1,1-Dichloroethane

Concen: 1.69 ppbv

RT: 5.95 min Scan# 469

Delta R.T. -0.01 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35


Tgt Ion: 63 Resp: 504098

Ion Ratio Lower Upper

63 100

98 5.6 2.9 8.7



100 4.1 1.7 5.1




Abundance Ion 63.10 (62.80 to 63.80): VL0

Ion 98.00 (97.70 to 98.70): VL0

Ion 100.00 (99.70 to 100.70): VL0





#25

Ethyl Acetate

Concen: 1.95 ppbv

RT: 6.71 min Scan# 595

Delta R.T. -0.00 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC002

Tgt Ion: 43 Resp: 764019


Ion Ratio Lower Upper

43 100

45 9.6 7.8 11.8

61 9.5 7.9 11.9

70 8.2 6.9 10.3






Abundance Ion 43.10 (42.80 to 43.80): VL0

Ion 45.10 (44.80 to 45.80): VL0

Ion 61.10 (60.80 to 61.80): VL0

Ion 70.10 (69.80 to 70.80): VL0



#26

Hexane

Concen: 2.07 ppbv

RT: 6.70 min Scan# 592

Delta R.T. -0.01 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35

Tgt Ion: 57 Resp: 404156


Ion Ratio Lower Upper

57 100

41 81.8 63.3 94.9

43 191.3 151.4 227.2

56 53.0 42.2 63.2






Abundance Ion 57.10 (56.80 to 57.80): VL0

Ion 41.10 (40.80 to 41.80): VL0

Ion 43.10 (42.80 to 43.80): VL0

Ion 56.10 (55.80 to 56.80): VL0





#27

Carbon Disulfide

Concen: 1.78 ppbv

RT: 5.43 min Scan# 384

Delta R.T. -0.00 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC002


Tgt Ion: 76 Resp: 691528

Ion Ratio Lower Upper




76 100

44 14.7 12.1 18.1

78 9.7 7.7 11.5



Abundance Ion 76.00 (75.70 to 76.70): VL0
Ion 44.10 (43.80 to 44.80): VL0
Ion 78.00 (77.70 to 78.70): VL0



#28

Methyl tert-Butyl Ether

Concen: 1.48 ppbv

RT: 6.00 min Scan# 478

Delta R.T. 0.01 min

Lab File: VL026824.D


Acq: 9 Feb 2016 17:35

Tgt Ion: 73 Resp: 645973



Ion Ratio Lower Upper


73 100

57 21.1 17.0 25.6



Abundance Ion 73.10 (72.80 to 73.80): VL0
Ion 57.10 (56.80 to 57.80): VL0









#29
Chloroform
Concen: 1.39 ppbv
RT: 6.79 min Scan# 607
Delta R.T. -0.01 min
Lab File: VL026824.D
Acq: 9 Feb 2016 17:35




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC002


Tgt Ion: 83 Resp: 608513
Ion Ratio Lower Upper
83 100
85 63.7 51.5 77.3



#30
Cyclohexane
Concen: 1.63 ppbv
RT: 8.32 min Scan# 859
Delta R.T. -0.01 min
Lab File: VL026824.D
Acq: 9 Feb 2016 17:35

Tgt Ion: 84 Resp: 328928
Ion Ratio Lower Upper
84 100
56 147.1 99.4 149.2
41 70.6 56.7 85.1






#31




cis-1,2-Dichloroethene
 Concen: 1.56 ppbv
 RT: 6.56 min Scan# 569
 Delta R.T. -0.01 min
 Lab File: VL026824.D
 Acq: 9 Feb 2016 17:35

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC002

Tgt Ion:	61	Resp:	332078
Ion	Ratio	Lower	Upper
61	100		
96	69.4	55.2	82.8
98	43.6	35.2	52.8




Abundance Ion 61.10 (60.80 to 61.80): VL0
 Ion 96.00 (95.70 to 96.70): VL0
 Ion 98.00 (97.70 to 98.70): VL0





#32


1,1,1-Trichloroethane
 Concen: 1.21 ppbv
 RT: 7.63 min Scan# 746
 Delta R.T. -0.01 min
 Lab File: VL026824.D
 Acq: 9 Feb 2016 17:35

Tgt Ion:	97	Resp:	603598
Ion	Ratio	Lower	Upper
97	100		
99	65.0	52.0	78.0
61	45.9	36.2	54.4



Abundance Ion 97.00 (96.70 to 97.70): VL0
 Ion 99.00 (98.70 to 99.70): VL0
 Ion 61.10 (60.80 to 61.80): VL0









#33
 1,4-Difluorobenzene
 Concen: 10.00 ppbv
 RT: 8.37 min Scan# 866
 Delta R.T. -0.01 min
 Lab File: VL026824.D
 Acq: 9 Feb 2016 17:35




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC002


Tgt Ion:114 Resp: 4564495
 Ion Ratio Lower Upper
 114 100
 63 24.3 19.7 29.5
 88 17.2 13.8 20.6



#34
 2-Butanone
 Concen: 2.20 ppbv
 RT: 6.24 min Scan# 518
 Delta R.T. -0.00 min
 Lab File: VL026824.D
 Acq: 9 Feb 2016 17:35

Tgt Ion: 43 Resp: 466773
 Ion Ratio Lower Upper
 43 100
 72 25.6 19.8 29.8





#35

Carbon Tetrachloride

Concen: 1.43 ppbv

RT: 8.20 min Scan# 839

Delta R.T. -0.01 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC002


Tgt Ion:117 Resp: 646799

Ion Ratio Lower Upper

117 100

119 97.9 76.6 115.0




121 31.6 25.0 37.4



Abundance Ion 117.00 (116.70 to 117.70): V

Ion 119.00 (118.70 to 119.70): V

Ion 121.00 (120.70 to 121.70): V



#36

Benzene

Concen: 1.93 ppbv

RT: 8.06 min Scan# 816

Delta R.T. -0.01 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35


Tgt Ion: 78 Resp: 737666

Ion Ratio Lower Upper

78 100

77 23.1 19.0 28.6



50 17.9 14.2 21.4




Abundance Ion 78.10 (77.80 to 78.80): VL0

Ion 77.10 (76.80 to 77.80): VL0

Ion 50.10 (49.80 to 50.80): VL0









#37
 1,2-Dichloroethane
 Concen: 1.38 ppbv
 RT: 7.41 min Scan# 709
 Delta R.T. -0.01 min
 Lab File: VL026824.D
 Acq: 9 Feb 2016 17:35




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC002


Tgt Ion: 62 Resp: 453037
 Ion Ratio Lower Upper
 62 100
 98 6.8 5.4 8.2



#38
 Trichloroethene
 Concen: 1.82 ppbv
 RT: 9.11 min Scan# 988
 Delta R.T. -0.01 min
 Lab File: VL026824.D
 Acq: 9 Feb 2016 17:35

Tgt Ion: 130 Resp: 329237
 Ion Ratio Lower Upper
 130 100
 95 107.6 83.0 124.6
 132 95.9 76.0 114.0
 97 69.1 54.0 81.0





#39

1,2-Dichloropropane

Concen: 2.13 ppbv

RT: 8.87 min Scan# 948

Delta R.T. -0.01 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC002


Tgt Ion: 63 Resp: 273942

Ion Ratio Lower Upper

63 100

41 67.8 53.2 79.8




62 70.4 56.0 84.0



Abundance Ion 63.10 (62.80 to 63.80): VL0

Ion 41.10 (40.80 to 41.80): VL0

Ion 62.10 (61.80 to 62.80): VL0



#40

1,4-Dioxane

Concen: 1.69 ppbv

RT: 9.20 min Scan# 1002

Delta R.T. 0.03 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35

Tgt Ion: 88 Resp: 53053


Ion Ratio Lower Upper

88 100

58 60.1 135.6 203.4#

43 0.0 332.3 498.5#

57 0.0 1660.5 2490.7#






Abundance Ion 88.10 (87.80 to 88.80): VL0

Ion 58.10 (57.80 to 58.80): VL0

Ion 43.10 (42.80 to 43.80): VL0

Ion 57.10 (56.80 to 57.80): VL0





#41

Tetrahydrofuran

Concen: 2.38 ppbv

RT: 7.18 min Scan# 671

Delta R.T. 0.01 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC002


Tgt Ion: 42 Resp: 228677

Ion Ratio Lower Upper

42 100

72 40.9 33.4 50.0




71 39.2 32.1 48.1



Abundance Ion 42.10 (41.80 to 42.80): VL0

Ion 72.10 (71.80 to 72.80): VL0

Ion 71.10 (70.80 to 71.80): VL0



#42

Bromodichloromethane

Concen: 1.43 ppbv

RT: 9.07 min Scan# 981

Delta R.T. -0.01 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35


Tgt Ion: 83 Resp: 609722

Ion Ratio Lower Upper

83 100

85 61.9 51.8 77.6



127 7.5 6.4 9.6




Abundance Ion 83.00 (82.70 to 83.70): VL0

Ion 85.00 (84.70 to 85.70): VL0

Ion 126.90 (126.60 to 127.60): V









#43
Methyl Methacrylate
Concen: 1.74 ppbv
RT: 9.31 min Scan# 1021
Delta R.T. -0.00 min
Lab File: VL026824.D
Acq: 9 Feb 2016 17:35




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC002


Tgt Ion: 69 Resp: 256454
Ion Ratio Lower Upper
69 100
41 133.9 106.5 159.7
100 34.2 27.2 40.8



#44
2,2,4-Trimethylpentane
Concen: 2.15 ppbv
RT: 9.13 min Scan# 991
Delta R.T. -0.01 min
Lab File: VL026824.D
Acq: 9 Feb 2016 17:35

Tgt Ion: 57 Resp: 1306104
Ion Ratio Lower Upper
57 100
41 29.4 23.2 34.8
56 32.6 26.2 39.4









#45
t-1,3-Dichloropropene
Concen: 1.32 ppbv
RT: 10.69 min Scan# 1248
Delta R.T. -0.01 min
Lab File: VL026824.D
Acq: 9 Feb 2016 17:35




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC002


Tgt Ion: 75 Resp: 335677
Ion Ratio Lower Upper
75 100
77 30.8 25.0 37.6



#46
cis-1,3-Dichloropropene
Concen: 1.53 ppbv
RT: 10.06 min Scan# 1144
Delta R.T. -0.01 min
Lab File: VL026824.D
Acq: 9 Feb 2016 17:35

Tgt Ion: 75 Resp: 417660
Ion Ratio Lower Upper
75 100
39 56.4 45.0 67.6
77 30.4 25.2 37.8









#47
 1,1,2-Trichloroethane
 Concen: 1.45 ppbv
 RT: 10.92 min Scan# 1285
 Delta R.T. -0.01 min
 Lab File: VL026824.D
 Acq: 9 Feb 2016 17:35




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC002


Tgt Ion: 97 Resp: 295716
 Ion Ratio Lower Upper
 97 100
 83 88.2 69.7 104.5
 85 55.5 45.3 67.9
 99 63.1 49.9 74.9



#48
 Dibromochloromethane
 Concen: 1.25 ppbv
 RT: 11.85 min Scan# 1437
 Delta R.T. -0.01 min
 Lab File: VL026824.D
 Acq: 9 Feb 2016 17:35

Tgt Ion: 129 Resp: 478982
 Ion Ratio Lower Upper
 129 100
 127 79.0 39.6 119.0





#49

Bromoform

Concen: 1.06 ppbv

RT: 14.86 min Scan# 1931

Delta R.T. -0.01 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC002


Tgt Ion: 173 Resp: 349775

Ion Ratio Lower Upper

173 100

175 48.2 39.2 58.8




171 52.4 41.4 62.0



Abundance Ion 172.90 (172.60 to 173.60): V

Ion 174.90 (174.60 to 175.60): V

Ion 170.90 (170.60 to 171.60): V



#50

4-Methyl-2-Pentanone

Concen: 1.87 ppbv

RT: 10.14 min Scan# 1157

Delta R.T. -0.00 min

Lab File: VL026824.D


Acq: 9 Feb 2016 17:35

Tgt Ion: 43 Resp: 552136

Ion Ratio Lower Upper



43 100


58 37.1 30.3 45.5



Abundance Ion 43.10 (42.80 to 43.80): VL0

Ion 58.10 (57.80 to 58.80): VL0









#51
2-Hexanone
Concen: 1.61 ppbv
RT: 11.68 min Scan# 1409
Delta R.T. 0.01 min
Lab File: VL026824.D
Acq: 9 Feb 2016 17:35




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC002


Tgt Ion: 43 Resp: 458462
Ion Ratio Lower Upper
43 100
58 50.1 40.4 60.6
98 0.6 0.2 0.4#



#52
Tetrachloroethene
Concen: 1.30 ppbv
RT: 12.84 min Scan# 1600
Delta R.T. -0.01 min
Lab File: VL026824.D
Acq: 9 Feb 2016 17:35

Tgt Ion: 164 Resp: 279026
Ion Ratio Lower Upper
164 100
166 127.6 100.3 150.5
129 100.1 80.0 120.0
131 96.9 76.7 115.1





#53

Toluene

Concen: 1.50 ppbv

RT: 11.28 min Scan# 1344

Delta R.T. -0.01 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35

Instrument :

MSVOA_L

ClientSampleId :





VSTDIC002

Tgt Ion: 91 Resp: 845070

Ion Ratio Lower Upper

91 100

92 58.3 47.3 70.9



#54

1,2-Dibromoethane

Concen: 1.29 ppbv

RT: 12.19 min Scan# 1493

Delta R.T. -0.01 min

Lab File: VL026824.D




Acq: 9 Feb 2016 17:35


Tgt Ion: 107 Resp: 444235

Ion Ratio Lower Upper

107 100

109 95.1 75.0 112.4









#55
Chlorobenzene-d5
Concen: 10.00 ppbv
RT: 13.79 min Scan# 1756
Delta R.T. -0.01 min
Lab File: VL026824.D
Acq: 9 Feb 2016 17:35




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC002


Tgt Ion:117 Resp: 4469471
Ion Ratio Lower Upper
117 100
82 60.7 51.9 77.9
119 0.0 48.5 72.7#



#56
1,1,1,2-Tetrachloroethane
Concen: 1.57 ppbv
RT: 13.83 min Scan# 1763
Delta R.T. -0.01 min
Lab File: VL026824.D
Acq: 9 Feb 2016 17:35

Tgt Ion:131 Resp: 370079
Ion Ratio Lower Upper
131 100
133 96.1 76.4 114.6
119 434.7 99.8 149.8#









#57
Chlorobenzene
Concen: 1.55 ppbv
RT: 13.86 min Scan# 1767
Delta R.T. -0.01 min
Lab File: VL026824.D
Acq: 9 Feb 2016 17:35




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC002


Tgt Ion: 112 Resp: 702313
Ion Ratio Lower Upper
112 100
77 68.1 53.3 79.9
114 30.2 28.9 35.3



#58
Ethyl Benzene
Concen: 1.53 ppbv
RT: 14.46 min Scan# 1865
Delta R.T. -0.01 min
Lab File: VL026824.D
Acq: 9 Feb 2016 17:35

Tgt Ion: 91 Resp: 1120180
Ion Ratio Lower Upper
91 100
106 28.8 24.9 37.3









#59
m/p-Xylene
Concen: 1.54 ppbv
RT: 14.76 min Scan# 1914
Delta R.T. -0.02 min
Lab File: VL026824.D
Acq: 9 Feb 2016 17:35




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC002


Tgt Ion: 91 Resp: 985053
Ion Ratio Lower Upper
91 100
106 87.4 38.2 57.2#



#60
o-Xylene
Concen: 1.40 ppbv
RT: 15.54 min Scan# 2042
Delta R.T. -0.01 min
Lab File: VL026824.D
Acq: 9 Feb 2016 17:35

Tgt Ion: 91 Resp: 994282
Ion Ratio Lower Upper
91 100
106 43.2 22.3 66.8





#61

Styrene

Concen: 1.34 ppbv

RT: 15.36 min Scan# 2013

Delta R.T. -0.01 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC002


Tgt Ion:104 Resp: 342450

Ion Ratio Lower Upper

104 100

78 50.5 39.8 59.6




103 47.3 38.2 57.4



Abundance Ion 104.10 (103.80 to 104.80): V

Ion 78.10 (77.80 to 78.80): VL0

Ion 103.10 (102.80 to 103.80): V



#62

Isopropylbenzene

Concen: 1.31 ppbv

RT: 16.59 min Scan# 2215

Delta R.T. -0.01 min

Lab File: VL026824.D


Acq: 9 Feb 2016 17:35

Tgt Ion:105 Resp: 1210981

Ion Ratio Lower Upper



105 100


120 25.2 21.0 31.4



Abundance Ion 105.10 (104.80 to 105.80): V

Ion 120.20 (119.90 to 120.90): V









#63
 1,1,2,2-Tetrachloroethane
 Concen: 1.42 ppbv
 RT: 15.53 min Scan# 2041
 Delta R.T. -0.01 min
 Lab File: VL026824.D
 Acq: 9 Feb 2016 17:35




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC002


Tgt Ion: 83 Resp: 647479
 Ion Ratio Lower Upper
 83 100
 131 9.1 4.7 14.1
 85 64.9 32.3 96.9



#64
 n-propylbenzene
 Concen: 1.17 ppbv
 RT: 17.57 min Scan# 2376
 Delta R.T. -0.01 min
 Lab File: VL026824.D
 Acq: 9 Feb 2016 17:35

Tgt Ion: 120 Resp: 302101
 Ion Ratio Lower Upper
 120 100
 91 500.2 363.0 544.6









#65
 tert-Butylbenzene
 Concen: 1.16 ppbv
 RT: 18.86 min Scan# 2587
 Delta R.T. -0.01 min
 Lab File: VL026824.D
 Acq: 9 Feb 2016 17:35




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC002


Tgt Ion: 119 Resp: 1088256
 Ion Ratio Lower Upper
 119 100
 91 77.7 61.5 92.3
 134 22.1 17.9 26.9



#66
 Benzyl Chloride
 Concen: 0.91 ppbv
 RT: 19.16 min Scan# 2636
 Delta R.T. -0.01 min
 Lab File: VL026824.D
 Acq: 9 Feb 2016 17:35

Tgt Ion: 91 Resp: 320386
 Ion Ratio Lower Upper
 91 100
 126 17.4 14.6 21.8
 128 5.8 4.6 6.8









#67
 sec-Butylbenzene
 Concen: 1.07 ppbv
 RT: 19.46 min Scan# 2685
 Delta R.T. -0.01 min
 Lab File: VL026824.D
 Acq: 9 Feb 2016 17:35




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC002


Tgt Ion: 105 Resp: 1467983
 Ion Ratio Lower Upper
 105 100
 134 19.1 16.1 24.1



#68
 1-Bromo-4-Fluorobenzene
 Concen: 8.38 ppbv
 RT: 16.33 min Scan# 2173
 Delta R.T. -0.01 min
 Lab File: VL026824.D
 Acq: 9 Feb 2016 17:35

Tgt Ion: 95 Resp: 3177756
 Ion Ratio Lower Upper
 95 100
 174 69.3 55.6 83.4
 175 4.9 3.9 5.9









#69
p-Isopropyltoluene
Concen: 1.07 ppbv
RT: 19.84 min Scan# 2748
Delta R.T. -0.01 min
Lab File: VL026824.D
Acq: 9 Feb 2016 17:35




Instrument :
MSVOA_L
ClientSampleId :
VSTDIC002


Tgt Ion: 119 Resp: 1178447
Ion Ratio Lower Upper
119 100
134 25.6 21.3 31.9
91 26.9 21.5 32.3



#70
n-Butylbenzene
Concen: 1.00 ppbv
RT: 20.80 min Scan# 2906
Delta R.T. -0.01 min
Lab File: VL026824.D
Acq: 9 Feb 2016 17:35

Tgt Ion: 91 Resp: 1149510
Ion Ratio Lower Upper
91 100
92 53.2 44.2 66.4
134 22.8 19.3 28.9









#71
 2-Chlorotoluene
 Concen: 1.17 ppbv
 RT: 17.47 min Scan# 2360
 Delta R.T. -0.01 min
 Lab File: VL026824.D
 Acq: 9 Feb 2016 17:35




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC002


Tgt Ion: 91 Resp: 844009
 Ion Ratio Lower Upper
 91 100
 126 30.6 12.2 48.8



#72
 4-Ethyltoluene
 Concen: 1.17 ppbv
 RT: 17.87 min Scan# 2425
 Delta R.T. -0.01 min
 Lab File: VL026824.D
 Acq: 9 Feb 2016 17:35

Tgt Ion: 105 Resp: 943701
 Ion Ratio Lower Upper
 105 100
 120 28.9 23.8 35.6
 91 13.3 10.2 15.4
 77 13.4 10.8 16.2









#73
 1,3,5-Trimethylbenzene
 Concen: 1.13 ppbv
 RT: 18.03 min Scan# 2452
 Delta R.T. -0.01 min
 Lab File: VL026824.D
 Acq: 9 Feb 2016 17:35




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC002


Tgt Ion:105 Resp: 894797
 Ion Ratio Lower Upper
 105 100
 120 45.3 39.4 59.0



#74
 1,2,4-Trimethylbenzene
 Concen: 1.14 ppbv
 RT: 18.88 min Scan# 2590
 Delta R.T. -0.01 min
 Lab File: VL026824.D
 Acq: 9 Feb 2016 17:35

Tgt Ion:105 Resp: 1002549
 Ion Ratio Lower Upper
 105 100
 120 49.3 41.3 61.9
 91 56.0 43.3 64.9
 77 23.3 18.3 27.5









#75
 1,3-Dichlorobenzene
 Concen: 1.04 ppbv
 RT: 19.17 min Scan# 2638
 Delta R.T. -0.01 min
 Lab File: VL026824.D
 Acq: 9 Feb 2016 17:35




Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC002


Tgt Ion:146 Resp: 597522
 Ion Ratio Lower Upper
 146 100
 111 43.2 22.1 66.3
 148 63.4 32.2 96.6



#76
 1,4-Dichlorobenzene
 Concen: 1.01 ppbv
 RT: 19.32 min Scan# 2663
 Delta R.T. -0.01 min
 Lab File: VL026824.D
 Acq: 9 Feb 2016 17:35

Tgt Ion:146 Resp: 580887
 Ion Ratio Lower Upper
 146 100
 111 43.2 21.6 64.8
 148 65.4 32.2 96.6









#77
 1,2-Dichlorobenzene
 Concen: 0.95 ppbv
 RT: 20.08 min Scan# 2787
 Delta R.T. -0.01 min
 Lab File: VL026824.D
 Acq: 9 Feb 2016 17:35

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC002

Tgt Ion	Ratio	Lower	Upper
146	100		
111	46.1	23.1	69.3
148	64.0	32.1	96.5




Abundance Ion 146.00 (145.70 to 146.70): V
 Ion 111.10 (110.80 to 111.80): V
 Ion 148.00 (147.70 to 148.70): V






#78
 Hexachloro-1,3-Butadiene
 Concen: 0.95 ppbv
 RT: 24.72 min Scan# 3548
 Delta R.T. -0.00 min
 Lab File: VL026824.D
 Acq: 9 Feb 2016 17:35

Tgt Ion	Ratio	Lower	Upper
225	100		
223	62.5	50.2	75.2
227	64.3	52.1	78.1



Abundance Ion 224.90 (224.60 to 225.60): V
 Ion 222.90 (222.60 to 223.60): V
 Ion 226.90 (226.60 to 227.60): V





#79

Naphthalene

Concen: 0.70 ppbv

RT: 23.98 min Scan# 3427

Delta R.T. -0.00 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC002


Tgt Ion:128 Resp: 530930

Ion Ratio Lower Upper

128 100

127 12.3 10.0 15.0




129 10.6 8.7 13.1



Abundance Ion 128.10 (127.80 to 128.80): V

Ion 127.10 (126.80 to 127.80): V

Ion 129.10 (128.80 to 129.80): V



#80

Naphthalene, 2-methyl-

Concen: 1.60 ppbv

RT: 26.09 min Scan# 3774

Delta R.T. 0.01 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35


Tgt Ion:142 Resp: 210417

Ion Ratio Lower Upper

142 100

141 86.8 67.2 100.8



115 35.7 27.5 41.3




Abundance Ion 142.00 (141.70 to 142.70): V

Ion 141.00 (140.70 to 141.70): V

Ion 115.00 (114.70 to 115.70): V





#81

1,2,4-Trichlorobenzene

Concen: 0.81 ppbv

RT: 23.78 min Scan# 3394

Delta R.T. -0.00 min

Lab File: VL026824.D

Acq: 9 Feb 2016 17:35

Instrument :

MSVOA_L

ClientSampleId :

VSTDIC002


Tgt Ion:180 Resp: 271057

Ion Ratio Lower Upper

180 100

182 96.9 77.3 115.9

184 30.0 24.5 36.7



Abundance Ion 180.00 (179.70 to 180.70): V

Ion 182.00 (181.70 to 182.70): V

Ion 184.00 (183.70 to 184.70): V

