

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL020916\
 Data File : VL026851.D
 Acq On : 10 Feb 2016 13:42
 Operator : SY/AP
 Sample : H1388-02DL 20X
 Misc : 400mL/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IAS-2DL

Quant Time: Feb 11 03:06:52 2016
 Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL020916AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Feb 10 03:14:04 2016
 QLast Update : Wed Feb 10 03:14:04 2016
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	6.68	49	1608162	10.00	ppbv	-0.01
33) 1,4-Difluorobenzene	8.36	114	4709758	10.00	ppbv	-0.02
55) Chlorobenzene-d5	13.79	117	4606388	10.00	ppbv	-0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	16.33	95	3275946	9.91	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Chlorodifluoromethane	3.60	51	564964	2.17	ppbv	98
13) Ethanol	4.36	45	326115	26.90	ppbv	97
15) Acetone	4.63	43	1769451	8.01	ppbv	99
19) Isopropyl Alcohol	4.81	45	152291	1.93	ppbv #	1

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : W:\HPCHEM1\MSVOA L\DATA\VL020916\
Data File : VL026851.D
Acq On : 10 Feb 2016 13:42
Operator : SY/AP
Sample : H1388-02DL 20X
Misc : 400mL/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IAS-2DL

Quant Time: Feb 11 03:06:52 2016
Quant Method : W:\HPCHEM1\MSVOA L\METHODS\VL020916AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Feb 10 03:14:04 2016
QLast Update : Wed Feb 10 03:14:04 2016
Response via : Initial Calibration

