

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL021021\
 Data File : VL036335.D
 Acq On : 11 Feb 2021 2:33
 Operator : SY/AP
 Sample : M1343-05DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SR-5BDL

Quant Time: Feb 11 21:13:25 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL021021AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Feb 10 16:04:02 2021
 QLast Update : Wed Feb 10 16:04:02 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1676776	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.15	114	3658372	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.07	117	3329140	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.39	95	2767996	10.11	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Chlorodifluoromethane	2.97	51	148714	0.714	ppbv	92
9) Propene	3.01	41	23892	0.308	ppbv	89
13) Ethanol	3.59	45	2473496	169.533	ppbv	89
15) Acetone	3.84	43	195462	0.954	ppbv #	84
19) Isopropyl Alcohol	3.95	45	1394088m	12.277	ppbv	
36) Benzene	6.87	78	47367	0.152	ppbv #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL021021\
Data File : VL036335.D
Acq On : 11 Feb 2021 2:33
Operator : SY/AP
Sample : M1343-05DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SR-5BDL

Quant Time: Feb 11 21:13:25 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL021021AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Feb 10 16:04:02 2021
QLast Update : Wed Feb 10 16:04:02 2021
Response via : Initial Calibration

