

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL021021\
 Data File : VL036340.D
 Acq On : 11 Feb 2021 8:33
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

MMDadoda
 2/12/2021 1:42:53 PM

Quant Time: Feb 11 21:18:54 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL021021AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Feb 10 16:04:02 2021
 QLast Update : Wed Feb 10 16:04:02 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1615823	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.15	114	3507454	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.06	117	3292740	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.39	95	2631559	9.71	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.86	43	37300	0.189	ppbv	# 17
20) Methylene Chloride	4.33	84	10973m	0.172	ppbv	
35) Carbon Tetrachloride	6.99	117	15454m	0.066	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL021021\
Data File : VL036340.D
Acq On : 11 Feb 2021 8:33
Operator : SY/AP
Sample : VIBLK
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

MMDadoda
2/12/2021 1:42:53 PM

Quant Time: Feb 11 21:18:54 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL021021AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Feb
QLast Update : Wed Feb 10 16:04:02 2021
Response via : Initial Calibration

