

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL021021\
 Data File : VL036341.D
 Acq On : 11 Feb 2021 9:13
 Operator : SY/AP
 Sample : QC020921 CAN 10673
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC020921 CAN 10673

Quant Time: Feb 11 09:56:37 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL021021AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_LWed Feb 10 16:04:02 2021
 QLast Update : Wed Feb 10 16:04:02 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1654859	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	3482204	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.07	117	3276148	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.40	95	2786460	10.34	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL021021\
Data File : VL036341.D
Acq On : 11 Feb 2021 9:13
Operator : SY/AP
Sample : OC020921 CAN 10673
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC020921 CAN 10673

Quant Time: Feb 11 09:56:37 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL021021AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Feb 10 16:04:02 2021
QLast Update : Wed Feb 10 16:04:02 2021
Response via : Initial Calibration

