

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL021419\
 Data File : VL033223.D
 Acq On : 15 Feb 2019 9:07
 Operator : SY/AP
 Sample : VIBLK
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VIBLK

Manual Integrations
 APPROVED

MMDadoda
 2/21/2019 1:03:25 PM

Quant Time: Feb 21 01:56:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL020519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Feb 21 01:56:24 2019
 QLast Update : Wed Feb 20 06:00:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	737517	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	1633392	10.00	ppbv	-0.01
55) Chlorobenzene-d5	12.24	117	2412837	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	1537444	8.18	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	81.80%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.92	43	14936m	0.180	ppbv	
20) Methylene Chloride	4.40	84	5919m	0.179	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL021419\
Data File : VL033223.D
Acq On : 15 Feb 2019 9:07
Operator : SY/AP
Sample : VIBLK
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Manual Integrations
APPROVED

MMDadoda
2/21/2019 1:03:25 PM

Quant Time: Feb 21 01:56:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL020519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Feb
QLast Update : Wed Feb 20 06:00:52 2019
Response via : Initial Calibration

