

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL021419\
 Data File : VL033227.D
 Acq On : 15 Feb 2019 11:41
 Operator : SY/AP
 Sample : K1483-05DL 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 ZONE-EDL

Quant Time: Feb 21 02:37:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL020519AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Feb 20 06:00:52 2019
 QLast Update : Wed Feb 20 06:00:52 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	811843	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	1797814	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.25	117	1859967	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.58	95	1473290	10.16	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
7) Chloroethane	3.61	64	5134	0.291	ppbv	# 79
18) 1,1-Dichloroethene	4.35	96	7788	0.200	ppbv	95
24) 1,1-Dichloroethane	5.09	63	356428	3.179	ppbv	99
32) 1,1,1-Trichloroethane	6.62	97	1729575	10.617	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL021419\
Data File : VL033227.D
Acq On : 15 Feb 2019 11:41
Operator : SY/AP
Sample : K1483-05DL 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
ZONE-EDL

Quant Time: Feb 21 02:37:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL020519AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Feb 20 06:00:52 2019
QLast Update : Wed Feb 20 06:00:52 2019
Response via : Initial Calibration

