

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL021422\
 Data File : VL038340.D
 Acq On : 14 Feb 2022 14:24
 Operator : SY/AP
 Sample : QC020722 CAN 10448
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC020722 CAN 10448

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/15/2022
 Supervised By :Mahesh Dadoda 02/18/2022

Quant Time: Feb 15 00:40:32 2022

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL012422AIR.M

Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Mon Jan 24 14:12:48 2022

QLast Update : Mon Jan 24 14:12:48 2022

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.658	49	890224	10.000	ppbv	-0.02
33) 1,4-Difluorobenzene	7.160	114	2229891	10.000	ppbv	-0.02
55) Chlorobenzene-d5	12.069	117	1916276	10.000	ppbv	#-0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.397	95	1482575	10.193	ppbv	-0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.900%
Target Compounds						
20) Methylene Chloride	4.340	84	6897m	0.151	ppbv	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL021422\
Data File : VL038340.D
Acq On : 14 Feb 2022 14:24
Operator : SY/AP
Sample : QC020722 CAN 10448
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC020722 CAN 10448

Quant Time: Feb 15 00:40:32 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL012422AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Jan
QLast Update : Mon Jan 24 14:12:48 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 02/15/2022
Supervised By : Mahesh Dadoda 02/18/2022

