

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL021422\
 Data File : VL038346.D
 Acq On : 14 Feb 2022 18:49
 Operator : SY/AP
 Sample : N1442-01DL2 6000X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV1DL3

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/15/2022
 Supervised By :Mahesh Dadoda 02/18/2022

Quant Time: Feb 15 00:41:39 2022

Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL012422AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Jan 24 2022

QLast Update : Mon Jan 24 14:12:48 2022

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.648	49	908301	10.000	ppbv	-0.03
33) 1,4-Difluorobenzene	7.150	114	2327743	10.000	ppbv	-0.03
55) Chlorobenzene-d5	12.057	117	1997654	10.000	ppbv	#-0.03
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.384	95	1615449	10.654	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	106.500%
Target Compounds						
						Qvalue
15) Acetone	3.857	43	120038m	1.499	ppbv	
38) Trichloroethene	7.806	130	8387m	0.109	ppbv	
52) Tetrachloroethene	11.195	164	712742	10.044	ppbv	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL021422\
Data File : VL038346.D
Acq On : 14 Feb 2022 18:49
Operator : SY/AP
Sample : N1442-01DL2 6000X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV1DL3

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/15/2022
Supervised By :Mahesh Dadoda 02/18/2022

Quant Time: Feb 15 00:41:39 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL012422AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Jan
QLast Update : Mon Jan 24 14:12:48 2022
Response via : Initial Calibration

