

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL021523\
 Data File : VL040174.D
 Acq On : 15 Feb 2023 13:37
 Operator : SY/MD
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/16/2023
 Supervised By :Mahesh Dadoda 02/18/2023

Quant Time: Feb 15 14:07:31 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL021523AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Wed Feb 15 11:34:13 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.189	49	953199	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.652	114	2421043	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.475	117	2148819	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.786	95	1764622	9.881	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.800%
Target Compounds						
					Qvalue	
5) Vinyl Chloride	2.915	62	1596m	0.027	ppbv	
32) 1,1,1-Trichloroethane	5.986	97	4067m	0.028	ppbv	
35) Carbon Tetrachloride	6.488	117	3514m	0.026	ppbv	
38) Trichloroethene	7.285	130	2587m	0.025	ppbv	
52) Tetrachloroethene	10.629	164	1376m	0.023	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.069	83	2544m	0.034	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL021523\
Data File : VL040174.D
Acq On : 15 Feb 2023 13:37
Operator : SY/MD
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Quant Time: Feb 15 14:07:31 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL021523AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed Feb 15 11:34:13 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 02/16/2023
Supervised By : Mahesh Dadoda 02/18/2023

