

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL021524\
 Data File : VL041074.D
 Acq On : 15 Feb 2024 16:49
 Operator : SY/MD
 Sample : P1490-01DL 20X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV1DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/19/2024
 Supervised By :Mahesh Dadoda 02/20/2024

Quant Time: Feb 16 04:10:30 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL012324AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Fri Feb 16 04:05:22 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.192	49	1255368	10.000	ppbv	0.02
33) 1,4-Difluorobenzene	6.652	114	3405584	10.000	ppbv	0.01
55) Chlorobenzene-d5	11.471	117	2910608	10.000	ppbv	0.02
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.780	95	2257477	10.269	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.700%
Target Compounds						
52) Tetrachloroethene	10.674	164	33076230	354.120	ppbv	Qvalue 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL021524\
Data File : VL041074.D
Acq On : 15 Feb 2024 16:49
Operator : SY/MD
Sample : P1490-01DL 20X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV1DL

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 02/19/2024
Supervised By : Mahesh Dadoda 02/20/2024

Quant Time: Feb 16 04:10:30 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL012324AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Fri Feb 16 04:05:22 2024
Response via : Initial Calibration

