

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL021524\
 Data File : VL041077.D
 Acq On : 16 Feb 2024 9:18
 Operator : SY/MD
 Sample : P1490-01DL2 600X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 SV1DL2

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 02/19/2024
 Supervised By :Mahesh Dadoda 02/20/2024

Quant Time: Feb 17 00:13:00 2024

Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL012324AIR.M

Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 2

QLast Update : Sat Feb 17 00:11:10 2024

Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.189	49	1143995	10.000	ppbv	0.01
33) 1,4-Difluorobenzene	6.648	114	2925259	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.465	117	2602649	10.000	ppbv	0.01
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.777	95	1954681	9.943	ppbv	0.02
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.400%
Target Compounds						Qvalue
52) Tetrachloroethene	10.619	164	3195926	39.834	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL021524\
Data File : VL041077.D
Acq On : 16 Feb 2024 9:18
Operator : SY/MD
Sample : P1490-01DL2 600X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
SV1DL2

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 02/19/2024
Supervised By :Mahesh Dadoda 02/20/2024

Quant Time: Feb 17 00:13:00 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL012324AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Sat Feb 17 00:11:10 2024
Response via : Initial Calibration

