

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL021622\
Data File : VL038366.D
Acq On : 16 Feb 2022 14:38
Operator : SY/AP
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Feb 17 06:34:03 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL012422AIR.M
Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Mon Jan 24 14:12:48 2022
QLast Update : Mon Jan 24 14:12:48 2022
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.63	49	841778	10.00	ppbv	-0.04
33) 1,4-Difluorobenzene	7.13	114	2095978	10.00	ppbv	-0.05
55) Chlorobenzene-d5	12.04	117	1802358	10.00	ppbv	-0.05

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.37	95	1419619	10.38	ppbv	-0.05
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.80%

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL021622\
Data File : VL038366.D
Acq On : 16 Feb 2022 14:38
Operator : SY/AP
Sample : VIBLK
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VIBLK

Quant Time: Feb 17 06:34:03 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL012422AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Jan 24 14:12:48 2022
QLast Update : Mon Jan 24 14:12:48 2022
Response via : Initial Calibration

