

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022222\
 Data File : VL038383.D
 Acq On : 22 Feb 2022 9:56
 Operator : SY/AP
 Sample : VL0222ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0222ABL01

Quant Time: Feb 22 19:54:14 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL022222AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Feb 22 19:45:51 2022
 QLast Update : Tue Feb 22 19:45:51 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.645	49	972503	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	7.150	114	2370312	10.000	ppbv	0.00
55) Chlorobenzene-d5	12.053	117	2040678	10.000	ppbv	# 0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.388	95	1662747	9.848	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022222\
Data File : VL038383.D
Acq On : 22 Feb 2022 9:56
Operator : SY/AP
Sample : VL0222ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0222ABL01

Quant Time: Feb 22 19:54:14 2022
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL022222AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Feb 22 19:45:51 2022
QLast Update : Tue Feb 22 19:45:51 2022
Response via : Initial Calibration

