

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL022620\
 Data File : VL034680.D
 Acq On : 26 Feb 2020 5:00
 Operator : SY/AP
 Sample : VSTDIC0.1
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.1

Manual Integrations
 APPROVED

MMDadoda
 2/27/2020 5:19:27 PM

Quant Time: Feb 26 06:43:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL022620AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Feb 26 06:43:12 2020
 QLast Update : Wed Feb 26 03:37:46 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1112034	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	2279889	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.11	117	2230809	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.46	95	1857573	10.14	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.40%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.23	62	6275	0.083	ppbv	90
32) 1,1,1-Trichloroethane	6.52	97	14505	0.086	ppbv #	85
35) Carbon Tetrachloride	7.02	117	15016m	0.086	ppbv	
38) Trichloroethene	7.85	130	7671m	0.094	ppbv	
52) Tetrachloroethene	11.26	164	7187m	0.095	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.72	83	17918m	0.087	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL022620\
Data File : VL034680.D
Acq On : 26 Feb 2020 5:00
Operator : SY/AP
Sample : VSTDIC0.1
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.1

Manual Integrations
APPROVED

MMDadoda
2/27/2020 5:19:27 PM

Quant Time: Feb 26 06:43:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL022620AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Feb
QLast Update : Wed Feb 26 03:37:46 2020
Response via : Initial Calibration

