

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL022620\
 Data File : VL034681.D
 Acq On : 26 Feb 2020 5:37
 Operator : SY/AP
 Sample : VSTDIC0.03
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.03

Manual Integrations
 APPROVED

MMDadoda
 2/27/2020 5:19:31 PM

Quant Time: Feb 26 06:45:54 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL022620AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Feb 26 2020
 QLast Update : Wed Feb 26 03:37:46 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	1081878	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	2236118	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.12	117	2189369	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.45	95	1846885	10.28	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.80%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.23	62	2221	0.030	ppbv	# 71
32) 1,1,1-Trichloroethane	6.51	97	5413m	0.033	ppbv	
35) Carbon Tetrachloride	7.02	117	4974m	0.029	ppbv	
38) Trichloroethene	7.84	130	2591m	0.032	ppbv	
52) Tetrachloroethene	11.26	164	2860m	0.038	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.72	83	6720m	0.033	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL022620\
Data File : VL034681.D
Acq On : 26 Feb 2020 5:37
Operator : SY/AP
Sample : VSTDICC0.03
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Manual Integrations
APPROVED

MMDadoda
2/27/2020 5:19:31 PM

Quant Time: Feb 26 06:45:54 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL022620AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Feb
QLast Update : Wed Feb 26 03:37:46 2020
Response via : Initial Calibration

