

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL022620\
 Data File : VL034687.D
 Acq On : 26 Feb 2020 10:49
 Operator : SY/AP
 Sample : L1314-01DL 4X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 E051-1100DL

Quant Time: Feb 27 15:39:57 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL022620AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Wed Feb 26 06:58:15 2020
 QLast Update : Wed Feb 26 06:58:15 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1078260	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	2293107	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.11	117	2243058	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.45	95	1926272	10.47	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.70%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
6) Bromomethane	3.44	94	357142	9.539	ppbv	97
26) Hexane	5.67	57	911325	9.091	ppbv #	42
30) Cyclohexane	7.13	84	819849	10.553	ppbv	99
35) Carbon Tetrachloride	7.02	117	1210420	7.882	ppbv	97
47) 1,1,2-Trichloroethane	9.48	97	769235	9.516	ppbv	96
52) Tetrachloroethene	11.24	164	645301	8.889	ppbv	97
59) m/p-Xylene	12.99	91	1844543	7.793	ppbv	99
60) o-Xylene	13.72	91	1575028	6.706	ppbv	98
72) 4-Ethyltoluene	15.87	105	2264235	8.661	ppbv	99
75) 1,3-Dichlorobenzene	17.03	146	988544	6.166	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL022620\
Data File : VL034687.D
Acq On : 26 Feb 2020 10:49
Operator : SY/AP
Sample : L1314-01DL 4X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
E051-1100DL

Quant Time: Feb 27 15:39:57 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL022620AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Feb 26 06:58:15 2020
QLast Update : Wed Feb 26 06:58:15 2020
Response via : Initial Calibration

