

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL022620\
 Data File : VL034697.D
 Acq On : 26 Feb 2020 21:13
 Operator : SY/AP
 Sample : L1663-01DL 10X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 EP-104DL

Manual Integrations
 APPROVED

MMDadoda
 2/27/2020 5:20:07 PM

Quant Time: Feb 27 15:54:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL022620AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Feb 27 15:54:49 2020
 QLast Update : Wed Feb 26 06:58:15 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1084470	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.19	114	2260808	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.12	117	2229734	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.46	95	1858435	10.16	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	101.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.58	45	390074	25.240	ppbv	98
15) Acetone	3.83	43	348210	2.818	ppbv	98
17) tert-Butyl alcohol	4.28	59	81585	0.862	ppbv	99
19) Isopropyl Alcohol	3.95	45	731497	9.938	ppbv	99
36) Benzene	6.90	78	21634m	0.111	ppbv	
53) Toluene	9.83	91	157815	0.760	ppbv	100
59) m/p-Xylene	13.01	91	97596m	0.415	ppbv	
74) 1,2,4-Trimethylbenzene	16.80	105	38152m	0.138	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL022620\
Data File : VL034697.D
Acq On : 26 Feb 2020 21:13
Operator : SY/AP
Sample : L1663-01DL 10X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
EP-104DL

Manual Integrations
APPROVED

MMDadoda
2/27/2020 5:20:07 PM

Quant Time: Feb 27 15:54:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL022620AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Feb
QLast Update : Wed Feb 26 06:58:15 2020
Response via : Initial Calibration

