

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL022620\
 Data File : VL034698.D
 Acq On : 26 Feb 2020 21:51
 Operator : SY/AP
 Sample : L1663-01DL2 40X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 EP-104DL2

Manual Integrations
 APPROVED

MMDadoda
 2/27/2020 5:20:09 PM

Quant Time: Feb 27 08:38:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL022620AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Feb 27 08:38:42 2020
 QLast Update : Wed Feb 26 06:58:15 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	1019155	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.19	114	2113026	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.13	117	2075526	10.00	ppbv	0.00

System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	14.46	95	1739889	10.22	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	102.20%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
13) Ethanol	3.58	45	75202m	5.178	ppbv	
15) Acetone	3.84	43	77581	0.668	ppbv	98
17) tert-Butyl alcohol	4.29	59	14702m	0.165	ppbv	
19) Isopropyl Alcohol	3.96	45	136855	1.978	ppbv #	96
20) Methylene Chloride	4.33	84	13271	0.307	ppbv #	76
26) Hexane	5.68	57	9738	0.103	ppbv #	48
53) Toluene	9.84	91	18729	0.096	ppbv	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL022620\
Data File : VL034698.D
Acq On : 26 Feb 2020 21:51
Operator : SY/AP
Sample : L1663-01DL2 40X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampled :
EP-104DL2

Manual Integrations
APPROVED

MMDadoda
2/27/2020 5:20:09 PM

Quant Time: Feb 27 08:38:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL022620AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Feb
QLast Update : Wed Feb 26 06:58:15 2020
Response via : Initial Calibration

