


Data Path : Z:\voasrv\HPCHEM1\MSVOA L\Data\VL022620\
 Data File : VL034675.D
 Acq On : 26 Feb 2020 1:48
 Operator : SY/AP
 Sample : BFB
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL022620AIR.M
 Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Wed Feb 26 06:58:15 2020
 Last Update : Wed Feb 26 06:58:15 2020



AutoFind: Scans 3627, 3628, 3629; Background Corrected with Scan 3603

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	8	40	25.7	197184	PASS
75	95	30	66	58.2	447168	PASS
95	95	100	100	100.0	768106	PASS
96	95	5	9	7.0	53573	PASS
173	174	0.00	2	0.8	4186	PASS
174	95	50	120	64.8	497920	PASS
175	174	4	9	7.2	35960	PASS
176	174	93	101	97.2	484032	PASS
177	176	5	9	6.1	29608	PASS