

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022723\
 Data File : VL040230.D
 Acq On : 27 Feb 2023 11:15
 Operator : SY/MD
 Sample : VL0227ABL01
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0227ABL01

Quant Time: Feb 28 00:17:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL021523AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
 QLast Update : Tue Feb 28 00:12:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.182	49	936568	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	6.645	114	2574354	10.000	ppbv	0.00
55) Chlorobenzene-d5	11.465	117	2185823	10.000	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	13.777	95	1696302	9.438	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	94.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL022723\
Data File : VL040230.D
Acq On : 27 Feb 2023 11:15
Operator : SY/MD
Sample : VL0227ABL01
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0227ABL01

Quant Time: Feb 28 00:17:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL021523AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug 26 06:05:16 2022
QLast Update : Tue Feb 28 00:12:21 2023
Response via : Initial Calibration

