

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL030221\
 Data File : VL036425.D
 Acq On : 2 Mar 2021 11:22
 Operator : SY/AP
 Sample : VL0302ABL01
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0302ABL01

Manual Integrations
 APPROVED

MMDadoda
 3/3/2021 1:52:23 PM

Quant Time: Mar 03 03:35:47 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL021021AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Feb 10
 QLast Update : Wed Feb 10 16:04:02 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	1587966	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	3692756	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.07	117	3296342	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.40	95	2662373	9.82	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.20%

Target Compounds					Ovalue
15) Acetone	3.85	43	49043m	0.253	ppbv
20) Methylene Chloride	4.33	84	10813m	0.173	ppbv

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL030221\
Data File : VL036425.D
Acq On : 2 Mar 2021 11:22
Operator : SY/AP
Sample : VL0302ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0302ABL01

Manual Integrations
APPROVED

MMDadoda
3/3/2021 1:52:23 PM

Quant Time: Mar 03 03:35:47 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL021021AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Feb
QLast Update : Wed Feb 10 16:04:02 2021
Response via : Initial Calibration

