

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL030321\
 Data File : VL036445.D
 Acq On : 3 Mar 2021 19:13
 Operator : SY/AP
 Sample : OC030321 CAN 10448
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC030321 CAN 10448

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 03/04/2021
 Supervised By : Mahesh Dadoda 03/04/2021

Quant Time: Mar 04 03:21:37 2021
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL030321A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Wed Mar 03 14:18:24 2021
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.68	49	1625134	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.19	114	3674517	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.10	117	3353288	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.43	95	2691922	9.76	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.88	43	55558m	0.299	ppbv	
20) Methylene Chloride	4.36	84	33941m	0.542	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL030321\
Data File : VL036445.D
Acq On : 3 Mar 2021 19:13
Operator : SY/AP
Sample : OC030321 CAN 10448
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC030321 CAN 10448

Quant Time: Mar 04 03:21:37 2021
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL030321.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Wed Mar 03 14:18:24 2021
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 03/04/2021
Supervised By :Mahesh Dadoda 03/04/2021

