

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL030520\
 Data File : VL034796.D
 Acq On : 5 Mar 2020 15:09
 Operator : SY/AP
 Sample : L1704-06DL 20X
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA-1-4DL

Quant Time: Mar 06 12:04:46 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL030420AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_LWed Mar 04 06:36:39 2020
 QLast Update : Wed Mar 04 06:36:39 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	998432	10.00	ppbv	-0.03
33) 1,4-Difluorobenzene	7.18	114	2048777	10.00	ppbv	-0.03
55) Chlorobenzene-d5	12.11	117	2111077	10.00	ppbv	-0.04

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.44	95	1835640	10.43	ppbv	-0.04
Spiked Amount	10.000	Range	65 - 135	Recovery	=	104.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
9) Propene	2.98	41	48227	1.077	ppbv	# 70
13) Ethanol	3.57	45	117433	7.655	ppbv	97
15) Acetone	3.82	43	1210003	10.120	ppbv	100
23) Vinyl Acetate	5.05	43	1474398	9.196	ppbv	98
53) Toluene	9.82	91	577370	3.116	ppbv	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL030520\
Data File : VL034796.D
Acq On : 5 Mar 2020 15:09
Operator : SY/AP
Sample : L1704-06DL 20X
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA-1-4DL

Quant Time: Mar 06 12:04:46 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL030420AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Wed Mar 04 06:36:39 2020
QLast Update : Wed Mar 04 06:36:39 2020
Response via : Initial Calibration

