

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL030819\
 Data File : VL033260.D
 Acq On : 8 Mar 2019 15:18
 Operator : SY/AP
 Sample : VSTDIC0.1
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDIC0.1

Manual Integrations
 APPROVED

MMDadoda
 3/11/2019 10:10:16 AM

Quant Time: Mar 08 17:33:40 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL030819AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Mar 08 16:12:14 2019
 QLast Update : Fri Mar 08 16:12:14 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	1093678	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.28	114	2440502	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.23	117	2491285	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	2002964	10.60	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	106.00%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
5) Vinyl Chloride	3.29	62	5323m	0.074	ppbv	
32) 1,1,1-Trichloroethane	6.62	97	16317m	0.090	ppbv	
35) Carbon Tetrachloride	7.12	117	15530	0.084	ppbv	87
38) Trichloroethene	7.96	130	7273	0.084	ppbv #	67
52) Tetrachloroethene	11.37	164	7076m	0.083	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.83	83	17312m	0.087	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL030819\
Data File : VL033260.D
Acq On : 8 Mar 2019 15:18
Operator : SY/AP
Sample : VSTDIC0.1
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDIC0.1

Manual Integrations
APPROVED

MMDadoda
3/11/2019 10:10:16 AM

Quant Time: Mar 08 17:33:40 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL030819AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Mar
QLast Update : Fri Mar 08 16:12:14 2019
Response via : Initial Calibration

