

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL030819\
 Data File : VL033264.D
 Acq On : 8 Mar 2019 18:56
 Operator : SY/AP
 Sample : VL0308ABL01
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VL0308ABL01

Manual Integrations
 APPROVED

MMDadoda
 3/11/2019 10:10:22 AM

Quant Time: Mar 09 04:55:38 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL030819AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Sat Mar 09 04:55:38 2019
 QLast Update : Sat Mar 09 03:32:03 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.76	49	969194	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.29	114	2190236	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.24	117	2469836	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.57	95	1968640	9.93	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	99.30%

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
15) Acetone	3.93	43	14371m	0.125	ppbv	
20) Methylene Chloride	4.41	84	7273m	0.154	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL030819\
Data File : VL033264.D
Acq On : 8 Mar 2019 18:56
Operator : SY/AP
Sample : VL0308ABL01
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VL0308ABL01

Manual Integrations
APPROVED

MMDadoda
3/11/2019 10:10:22 AM

Quant Time: Mar 09 04:55:38 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL030819AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Sat Mar
QLast Update : Sat Mar 09 03:32:03 2019
Response via : Initial Calibration

