

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL030822\
 Data File : VL038558.D
 Acq On : 8 Mar 2022 14:16
 Operator : SY/AP
 Sample : VSTDICC0.03
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 VSTDICC0.03

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/10/2022
 Supervised By :Mahesh Dadoda 03/14/2022

Quant Time: Mar 08 14:53:57 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL030822A
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument:
 QLast Update : Tue Mar 08 13:53:01 2022
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.66	49	952815	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	2355636	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.07	117	2051907	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.40	95	1570221	9.25	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	92.50%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
5) Vinyl Chloride	3.25	62	3918	0.06	ppbv	90
32) 1,1,1-Trichloroethane	6.51	97	9349m	0.06	ppbv	
35) Carbon Tetrachloride	7.01	117	10085	0.06	ppbv #	85
38) Trichloroethene	7.83	130	4942m	0.07	ppbv	
52) Tetrachloroethene	11.20	164	4893m	0.07	ppbv	
63) 1,1,2,2-Tetrachloroethane	13.67	83	12920m	0.07	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL030822\
Data File : VL038558.D
Acq On : 8 Mar 2022 14:16
Operator : SY/AP
Sample : VSTDICC0.03
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
VSTDICC0.03

Quant Time: Mar 08 14:53:57 2022
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL030822\
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Tue Mar 08 13:53:01 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 03/10/2022
Supervised By :Mahesh Dadoda 03/14/2022

