

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL030823\
 Data File : VL040257.D
 Acq On : 8 Mar 2023 16:53
 Operator : SY/MD
 Sample : 01836-03DL 20X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 DC-3723-SSDL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/09/2023
 Supervised By :Mahesh Dadoda 03/09/2023

Quant Time: Mar 09 04:14:21 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL021523AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Wed Mar 01 04:21:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.179	49	875149	10.000	ppbv	-0.02
33) 1,4-Difluorobenzene	6.639	114	2176564	10.000	ppbv	-0.03
55) Chlorobenzene-d5	11.455	117	1904325	10.000	ppbv	-0.03
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.767	95	1511500	9.652	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.500%
Target Compounds						
					Qvalue	
13) Ethanol	3.253	45	6894m	1.757	ppbv	
15) Acetone	3.475	43	46893	0.445	ppbv	95
20) Methylene Chloride	3.925	84	44505	0.811	ppbv	91
26) Hexane	5.198	57	18156	0.216	ppbv #	32
52) Tetrachloroethene	10.606	164	168861	3.193	ppbv	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL030823\
Data File : VL040257.D
Acq On : 8 Mar 2023 16:53
Operator : SY/MD
Sample : 01836-03DL 20X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
DC-3723-SSDL

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 03/09/2023
Supervised By :Mahesh Dadoda 03/09/2023

Quant Time: Mar 09 04:14:21 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL021523AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed Mar 01 04:21:39 2023
Response via : Initial Calibration

