

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL030823\
 Data File : VL040261.D
 Acq On : 8 Mar 2023 19:27
 Operator : SY/MD
 Sample : 01836-04DL 20X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 LM-3723-SSDL

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 03/09/2023
 Supervised By : Mahesh Dadoda 03/09/2023

Quant Time: Mar 09 04:15:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL021523AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Wed Mar 01 04:21:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.176	49	841651	10.000	ppbv	-0.03
33) 1,4-Difluorobenzene	6.639	114	2123703	10.000	ppbv	-0.03
55) Chlorobenzene-d5	11.452	117	1865725	10.000	ppbv	-0.04
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.764	95	1472517	9.598	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.000%
Target Compounds						
					Qvalue	
13) Ethanol	3.250	45	18354m	4.865	ppbv	
15) Acetone	3.468	43	72730	0.718	ppbv	94
20) Methylene Chloride	3.922	84	83392	1.581	ppbv	93
26) Hexane	5.198	57	41299	0.511	ppbv #	35
31) cis-1,2-Dichloroethene	5.066	61	7437m	0.097	ppbv	
38) Trichloroethene	7.282	130	7931m	0.087	ppbv	
52) Tetrachloroethene	10.603	164	317696m	6.156	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL030823\
Data File : VL040261.D
Acq On : 8 Mar 2023 19:27
Operator : SY/MD
Sample : 01836-04DL 20X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
LM-3723-SSDL

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 03/09/2023
Supervised By :Mahesh Dadoda 03/09/2023

Quant Time: Mar 09 04:15:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL021523AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed Mar 01 04:21:39 2023
Response via : Initial Calibration

