

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL030823\
 Data File : VL040262.D
 Acq On : 8 Mar 2023 20:05
 Operator : SY/MD
 Sample : 01836-04DL 40X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 LM-3723-SSDL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/09/2023
 Supervised By :Mahesh Dadoda 03/09/2023

Quant Time: Mar 09 04:16:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL021523AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Wed Mar 01 04:21:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	5.179	49	851759	10.000	ppbv	-0.02
33) 1,4-Difluorobenzene	6.639	114	2062444	10.000	ppbv	-0.03
55) Chlorobenzene-d5	11.452	117	1857969	10.000	ppbv	-0.04
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	13.764	95	1456548	9.534	ppbv	-0.03
Spiked Amount	10.000	Range	65 - 135	Recovery	=	95.300%
Target Compounds						
					Qvalue	
9) Propene	2.694	41	3559	0.086	ppbv #	76
13) Ethanol	3.256	45	8847m	2.317	ppbv	
15) Acetone	3.481	43	37615m	0.367	ppbv	
20) Methylene Chloride	3.922	84	47752	0.895	ppbv	98
26) Hexane	5.202	57	19334m	0.236	ppbv	
52) Tetrachloroethene	10.606	164	97225m	1.940	ppbv	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL030823\
Data File : VL040262.D
Acq On : 8 Mar 2023 20:05
Operator : SY/MD
Sample : 01836-04DL 40X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
LM-3723-SSDL

Quant Time: Mar 09 04:16:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL021523AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed Mar 01 04:21:39 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 03/09/2023
Supervised By :Mahesh Dadoda 03/09/2023

