

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL030920\
Data File : VL034855.D
Acq On : 10 Mar 2020 1:51
Operator : SY/AP
Sample : CAN 10294
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10294

Quant Time: Mar 11 09:26:31 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL030920AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar 09 16:42:53 2020
QLast Update : Mon Mar 09 16:42:53 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	873176	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.17	114	1978446	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.10	117	1914375	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.44	95	1537945	10.30	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.00%

Target Compounds	Ovalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL030920\
Data File : VL034855.D
Acq On : 10 Mar 2020 1:51
Operator : SY/AP
Sample : CAN 10294
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
CAN 10294

Quant Time: Mar 11 09:26:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL030920AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar 09 16:42:53 2020
QLast Update : Mon Mar 09 16:42:53 2020
Response via : Initial Calibration

