

Data Path : Z:\VOASRV\HPCHEM1\MSVOA L\DATA\VL030920\
 Data File : VL034857.D
 Acq On : 10 Mar 2020 3:06
 Operator : SY/AP
 Sample : L1161-14 0.03 LOQ
 Misc : 400ml/MSVOA L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 LOQ-AIR-01-QT1-2020

Manual Integrations
 APPROVED

MMDadoda
 3/11/2020 3:26:17 PM

Quant Time: Mar 11 05:14:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA L\METHODS\VL030920AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar 0
 QLast Update : Mon Mar 09 16:42:53 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	874738	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.18	114	1971055	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.11	117	1913874	10.00	ppbv	0.00

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.44	95	1538065	10.30	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	103.00%

Target Compounds					Ovalue
5) Vinyl Chloride	3.21	62	1861m	0.035	ppbv
32) 1,1,1-Trichloroethane	6.50	97	4410m	0.034	ppbv
38) Trichloroethene	7.83	130	2775m	0.042	ppbv
52) Tetrachloroethene	11.25	164	2321m	0.033	ppbv
63) 1,1,2,2-Tetrachloroethane	13.70	83	4623m	0.032	ppbv

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA L\DATA\VL030920\
Data File : VL034857.D
Acq On : 10 Mar 2020 3:06
Operator : SY/AP
Sample : L1161-14 0.03 LOQ
Misc : 400ml/MSVOA L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
LOQ-AIR-01-QT1-2020

Manual Integrations
APPROVED

MMDadoda
3/11/2020 3:26:17 PM

Quant Time: Mar 11 05:14:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA L\METHODS\VL030920AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Mon Mar
QLast Update : Mon Mar 09 16:42:53 2020
Response via : Initial Calibration

