

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL030922\
 Data File : VL038580.D
 Acq On : 9 Mar 2022 17:44
 Operator : SY/AP
 Sample : N1702-03DL2 40X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 IA-1DL2

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 03/10/2022
 Supervised By : Mahesh Dadoda 03/14/2022

Quant Time: Mar 10 01:06:50 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL030822A
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
 QLast Update : Tue Mar 08 21:18:44 2022
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.67	49	859047	10.00	ppbv	0.01
33) 1,4-Difluorobenzene	7.18	114	2152934	10.00	ppbv	0.01
55) Chlorobenzene-d5	12.09	117	1948421	10.00	ppbv	0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.41	95	1445364	9.76	ppbv	0.01
Spiked Amount	10.000	Range	65 - 135	Recovery	=	97.60%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
9) Propene	3.03	41	7490m	0.14	ppbv	
13) Ethanol	3.63	45	27779m	4.52	ppbv	
15) Acetone	3.86	43	398705	6.58	ppbv	95
20) Methylene Chloride	4.35	84	41165	1.13	ppbv	93
26) Hexane	5.69	57	231251	2.41	ppbv #	43
53) Toluene	9.81	91	322489	1.55	ppbv	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL030922\
Data File : VL038580.D
Acq On : 9 Mar 2022 17:44
Operator : SY/AP
Sample : N1702-03DL2 40X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
IA-1DL2

Quant Time: Mar 10 01:06:50 2022
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL030822
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument:
QLast Update : Tue Mar 08 21:18:44 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 03/10/2022
Supervised By :Mahesh Dadoda 03/14/2022

