

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_L\DATA\VL030922\
 Data File : VL038587.D
 Acq On : 10 Mar 2022 8:37
 Operator : SY/AP
 Sample : QC030922 CAN 10153
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 QC030922 CAN 10153

Quant Time: Mar 11 00:21:45 2022
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_L\METHODS\VL030822AIR.M
 Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Mar 08 21:18:44 2022
 QLast Update : Tue Mar 08 21:18:44 2022
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Bromochloromethane	5.65	49	773161	10.00	ppbv	0.00
33) 1,4-Difluorobenzene	7.16	114	2011016	10.00	ppbv	0.00
55) Chlorobenzene-d5	12.06	117	1779081	10.00	ppbv	-0.01

System Monitoring Compounds

68) 1-Bromo-4-Fluorobenzene	14.39	95	1299363	9.61	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	96.10%

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
20) Methylene Chloride	4.33	84	4066	0.12	ppbv	# 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_L\DATA\VL030922\
Data File : VL038587.D
Acq On : 10 Mar 2022 8:37
Operator : SY/AP
Sample : QC030922 CAN 10153
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
QC030922 CAN 10153

Quant Time: Mar 11 00:21:45 2022
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_L\METHODS\VL030822AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Tue Mar 08 21:18:44 2022
QLast Update : Tue Mar 08 21:18:44 2022
Response via : Initial Calibration

