

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL031025\
 Data File : VL042128.D
 Acq On : 10 Mar 2025 16:47
 Operator : SY/MD
 Sample : Q1532-04DL 10X
 Misc : 400mL/MSVOA_L
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_L
 ClientSampleId :
 4DL

Manual Integrations
 APPROVED

Reviewed By :Semsettin Yesilyurt 03/11/2025
 Supervised By :Mahesh Dadoda 03/11/2025

Quant Time: Mar 11 01:27:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M
 Quant Title : AIR ANALYSIS BY METHOD T0-15 Instrument: MSVOA_L Fri Aug 2
 QLast Update : Wed Feb 26 01:08:36 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	2.800	49	165173	10.000	ppbv	0.00
33) 1,4-Difluorobenzene	3.978	114	369716	10.000	ppbv	0.00
55) Chlorobenzene-d5	8.898	117	310848	10.000	ppbv	0.00
System Monitoring Compounds						
68) 1-Bromo-4-Fluorobenzene	10.390	95	235145	9.833	ppbv	0.00
Spiked Amount	10.000	Range	65 - 135	Recovery	=	98.300%
Target Compounds						
					Qvalue	
9) Propene	1.489	41	1802	0.184	ppbv	96
13) Ethanol	1.729	45	1793m	2.888	ppbv	
15) Acetone	1.842	43	10002	0.686	ppbv	92
20) Methylene Chloride	2.072	84	3747	0.679	ppbv #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_L\Data\VL031025\
Data File : VL042128.D
Acq On : 10 Mar 2025 16:47
Operator : SY/MD
Sample : Q1532-04DL 10X
Misc : 400mL/MSVOA_L
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_L
ClientSampleId :
4DL

Quant Time: Mar 11 01:27:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_L\methods\VL022525AIR.M
Quant Title : AIR ANALYSIS BY METHOD TO-15 Instrument: MSVOA_L Fri Aug
QLast Update : Wed Feb 26 01:08:36 2025
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : Semsettin Yesilyurt 03/11/2025
Supervised By : Mahesh Dadoda 03/11/2025

